

NOV 23 2011

Special 510(k) Submission – SDX-4343CS

Special 510(k) Summary

K112751

This summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of 21 CFR Part 807.92.

Date: September 15th, 2011

1. Company and Correspondent making the submission:

Name – Samsung Mobile Display Co., Ltd.
Address – San #24, Nongseo-Dong, Giheung-Gu, Yongin-Si,
Gyeonggi-Do, Korea, 446-711
Telephone – +82-18-323-4075
Fax – +82-31-209-4881
Mr. Kyung Hun Yoon / Senior Manager
Internet – <http://www.SAMSUNG.com>

2. Official Correspondent for this submission

Contact – Dave Kim
Address – 12946 Kimberley Ln, Houston, TX 77079
Tel : 713-467-2607
Fax : 713-464-8880
Email : davekim@mtech-inc.net

3. Device :

Trade/proprietary name(s) : SDX-4343CS
Common Name : Digital Flat Panel X-Ray Detector
Classification Name : Solid State X-ray Imaging Device

4. Predicate Device :

Manufacturer : Samsung Mobile Display Co., Ltd
Device(s) and 510K No. : SDX-4336CP (K012321, Decision Date – Feb 11, 2011)

5. Classifications Names & Citations :

21CFR 892.1650, MQB, Solid State X-ray Imaging Device, Class2

Samsung Mobile Display Co., Ltd.

p. 1 of 3

6. Description :

6.1 General

SDX-4343CS digital X-ray flat panel detectors can generate radiographic images of any part of the body. Each of them consists of a scintillator directly coupled to an a-SI TFT sensor. It makes high-resolution, high-sensitive digital images. SDX-4343CS is designed specifically to be integrated with an operating PC and a X-Ray generator to digitalize x-ray images into RAW files. The RAW files can be made to DICOM compatible image files for a radiographic diagnosis and analysis by the user's designated console software which is not included in SDX-4343CS, digital flat panel X-ray detector.

7. Indication for use :

SDX-4343CS Digital Flat Panel X-Ray Detector is indicated for digital imaging solution designed for providing general radiographic diagnosis of human anatomy for adults and pediatric care. It is intended to replace film based radiographic diagnostic systems and provide a case diagnosis and treatment planning for physicians and other health care professionals who are licensed by the law of the State in which he or she practices to use the device. Not to be used for mammography.

8. Comparison with predicate device :

Samsung Mobile Display Co., Ltd., believes that SDX-4343CS is substantially equivalent to the predicate devices; SDX-4336CP.

9. Safety, EMC and Performance Data :

Electrical, mechanical, environmental safety and performance testing according to standard EN/IEC 60601-1 was performed, and EMC testing was conducted in accordance with standard EN/IEC 60601-1-2(2001). All test results were satisfactory.

9. Conclusions :

In accordance with the Federal Food, Drug and Cosmetic Act, 21 CFR Part 807 and based on the information provided in this premarket notification, Samsung Mobile Display Co., Ltd. concludes that SDX-4343CS is safe and effective and substantially equivalent to its predicate device as described herein.

10. Samsung Mobile Display Co., Ltd. will update and include in this summary any other information deemed necessary by the FDA.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room – WO66-G609
Silver Spring, MD 20993-0002

Samsung Mobile Display Co., Ltd.
% Mr. Dave Kim
Medical Device Regulatory Affairs
Mtech Group
12946 Kimberly Lane
HOUSTON TX 77079

AUG 23 2013

Re: K112751

Trade/Device Name: Digital Flat Panel X-Ray Detector/SDX-4343CS
Regulation Number: 21 CFR 892.1680
Regulation Name: Stationary x-ray system
Regulatory Class: II
Product Code: MQB
Dated: November 16, 2011
Received: November 7, 2011

Dear Mr. Kim:

This letter corrects our substantially equivalent letter of November 23, 2011.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of

medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely Yours,



Janine M. Morris
Acting Director
Division of Radiological Devices
Office of In Vitro Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(K) Number (if known): K112751

Device Name: Digital Flat Panel X-Ray Detector /SDX-4343CS

Indications for Use:

SDX-4343CS Digital Flat Panel X-Ray Detector is indicated for digital imaging solution designed for providing general radiographic diagnosis of human anatomy for adults and pediatric care. It is intended to replace film based radiographic diagnostic systems and provide a case diagnosis and treatment planning for physicians and other health care professionals who are licensed by the law of the State in which he or she practices to use the device. Not to be used for mammography.

Prescription Use ☒
 (Part 21 CFR 801 Subpart D)


AND/OR

Over-The-Counter Use _____
 (Part 21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Page 1 of 1


(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety
510K K112751